## **BJ-3 DISINTEGRATION TESTER**

This instruments are used for detecting disintegration of solid in prescriptive conditions.

## **Applicable standards:**

- · National standard( Chinese codex ) /2010
- · Corporate standard(Dissolution tester )Q/12XQ0188-2010

## **Features:**

- There are eighteen tubes in three basket assembly.
- · Each test station is driven independently. Auto stop at timing point.
- The bath liquid can achieve an even temperature.
- · A MPU is used to automatically control the temperature, time and frequency of nacelle back and forth.
- · Timing at will. The preset and real time data can be displayed alternately in time.
- · Automation: auto-test, auto-diagnose, auto-alarm.

## **Specifications:**

- · Basket assembly :3
- · Nacelle back and forth frequency : $(30\sim32)$  /min
- · Nacelle back and forth range: (55±1)mm
- · Automatic time range : $(5\sim120)$ min
- · Accuracy of controlling temperature :(37±0.5)°C
- Power :220V/ 50Hz/ 600W or 110V/ 60Hz/ 600W
- Dimension :  $(540\times300\times420)$ mm<sup>3</sup>
- Weight :22kg

